

Airborne Contamination Identification Associates Ltd Certifies that

New Concept Technology 320 Busser Road PO Box 0297 Emigsville PA 17318-0297

"Medical Assembly Cleanroom" meets and exceeds ISO 14644-1 Class 6 on 0.3, 0.5 micron for airborne particulate cleanliness at rest.

"Medical Assembly Cleanroom" meets and exceeds ISO 14644-1 Class 6 on 0.3 0.5, 1.0 micron for airborne particulate cleanliness in production.

Certificate # 08-28-2025 Expires: September 01, 2026

Certified By:

Randall R. Leaman C.I.A.Q.P.

"Certified Indoor Air Quality Professional